

-----  
Midmark M11 - v1.0.5  
Total Cycles: 940

Name: RITTER2  
Sterilizer ID: V2313228  
04/01/2022 06:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.4 F 0.0  
2:00 86.9 F 0.0  
4:00 77.7 F 0.0  
6:00 107.3 F 0.3  
8:00 137.5 F 1.1  
10:00 173.5 F 2.7  
12:00 214.0 F 6.0  
14:00 211.9 F 9.8  
16:00 221.0 F 9.8  
18:00 230.0 F 9.5  
20:00 235.8 F 10.3  
22:00 240.3 F 12.2  
24:00 239.8 F 10.9  
26:00 252.8 F 17.5  
28:00 263.8 F 24.1  
30:00 271.1 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.5  
1:00 272.7 F 29.7  
2:00 272.8 F 30.2  
3:00 271.8 F 29.4  
4:00 272.8 F 30.0

Min 271.6 F 29.4  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.0 F 0.0  
5:00 227.6 F 0.0  
10:00 199.9 F 0.0  
15:00 204.6 F 0.0  
20:00 206.4 F 0.0  
25:00 210.9 F 0.0  
30:00 215.6 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 30:45  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:08:60  
CYCLE COMPLETE

04/01/2022 07:52 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_