
Midmark M11 - v1.0.5
Total Cycles: 114

Name: RITTER2
Sterilizer ID: V2313228
06/17/2021 01:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

C102: FILL MODE
STOP PRESSED
CHAMBER TEMP 164F
CHAMBER PRESS 0.0PSI
** ITEMS NOT STERILE **

06/17/2021 01:44 PM