

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001721  
9:50 27/03/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:13  
PRESSURIZING 3:19  
STERILIZING 4:25  
136.3 C 222kPa 4:25  
136.3 C 223kPa 4:55  
136.1 C 223kPa 5:25  
136.3 C 223kPa 5:55  
136.4 C 225kPa 6:25  
136.4 C 224kPa 6:55  
136.1 C 221kPa 7:25  
136.4 C 223kPa 7:55  
136.4 C 223kPa 7:55  
VENTING 7:55  
AIR DRYING 8:44  
CYCLE COMPLETE 9:31

-----  
Digital Signature #  
0D556CC7F0F62F41  
-----