
Midmark M11 - v1.0.5
Total Cycles: 694

Name: RITTER2
Sterilizer ID: V2313228
01/13/2022 10:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.5 F 0.0
2:00 149.2 F 2.3
4:00 168.8 F 5.8
6:00 191.3 F 10.6
8:00 214.9 F 5.0
10:00 225.7 F 10.6
12:00 229.4 F 7.7
14:00 240.3 F 11.6
16:00 244.7 F 13.3
18:00 257.3 F 20.2
20:00 268.7 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 272.1 F 29.6
2:00 272.6 F 30.0
3:00 272.3 F 30.3
4:00 271.9 F 29.5

Min 271.7 F 29.5
Max 272.6 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.9 F 0.0
5:00 216.7 F 0.0
10:00 218.2 F 0.0
15:00 219.2 F 0.0
20:00 225.0 F 0.0
25:00 225.3 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 21:05
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

01/13/2022 11:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____