
Midmark M11 - v1.0.0
Total Cycles: 1395

Name: RITTER 1
Sterilizer ID: V2178198
11/24/2021 02:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.9 F 0.0
2:00 155.0 F 1.6
4:00 158.5 F 4.2
6:00 188.5 F 7.5
8:00 202.7 F 11.8
10:00 215.7 F 5.4
12:00 227.4 F 9.7
14:00 236.0 F 10.3
16:00 238.9 F 11.1
18:00 235.8 F 8.7
20:00 249.1 F 15.0
22:00 260.1 F 21.1
24:00 270.2 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.4 F 29.4
2:00 272.4 F 29.0
3:00 271.3 F 28.5
4:00 271.9 F 29.3

Min 271.2 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.1 F 0.0
5:00 212.7 F 0.0
10:00 220.8 F 0.0
15:00 224.4 F 0.0
20:00 225.1 F 0.0
25:00 227.4 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:59
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:03:27
CYCLE COMPLETE

11/24/2021 03:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____