
Midmark M11 - v1.0.5
Total Cycles: 3064

Name: RITTER2
Sterilizer ID: V2313228
10/30/2023 12:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.3 F 0.0
2:00 171.9 F 3.0
4:00 200.5 F 8.2
6:00 217.2 F 13.4
8:00 227.3 F 8.9
10:00 234.9 F 9.6
12:00 242.9 F 13.3
14:00 243.9 F 12.9
16:00 257.6 F 20.2
18:00 269.6 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.6 F 29.8
2:00 271.5 F 29.1
3:00 272.7 F 29.9
4:00 271.9 F 29.3

Min 271.4 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.1 F 0.0
5:00 204.0 F 0.0
10:00 211.4 F 0.0
15:00 210.5 F 0.0
20:00 206.4 F 0.0
25:00 205.3 F 0.0
30:00 204.9 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 18:57
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

10/30/2023 01:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____