
Midmark M11 - v1.0.5
Total Cycles: 4345

Name: RITTER2
Sterilizer ID: V2313228
02/10/2025 01:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.3 F 0.0
2:00 143.8 F 1.9
4:00 154.8 F 4.9
6:00 184.6 F 9.3
8:00 214.5 F 7.8
10:00 226.2 F 7.6
12:00 237.9 F 12.1
14:00 234.5 F 8.7
16:00 242.4 F 11.3
18:00 245.0 F 12.8
20:00 256.3 F 19.0
22:00 266.9 F 25.5
24:00 272.9 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.4 F 29.6
2:00 272.0 F 29.8
3:00 272.2 F 29.4
4:00 272.3 F 30.0

Min 271.3 F 29.1
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.8 F 0.0
5:00 226.2 F 0.0
10:00 211.5 F 0.0
15:00 179.2 F 0.0
20:00 177.6 F 0.0
25:00 176.9 F 0.0
30:00 170.4 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:24
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:03:05
CYCLE COMPLETE

02/10/2025 02:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____