
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008390
7:28 10/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:52
STERILIZING 4:01
136.0 C 220kPa 4:01
136.3 C 221kPa 4:31
135.8 C 220kPa 5:01
136.1 C 221kPa 5:31
136.3 C 222kPa 6:01
136.4 C 223kPa 6:31
136.4 C 223kPa 7:01
136.2 C 222kPa 7:31
136.2 C 222kPa 7:31
VENTING 7:31
AIR DRYING 8:19
CYCLE COMPLETE 9:36

Digital Signature #
9F5B0DCB89F8E0EF
