
Midmark M11 - v1.0.5
Total Cycles: 1574

Name: RITTER2
Sterilizer ID: V2313228
10/06/2022 10:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.5 F 0.0
2:00 200.5 F 0.8
4:00 215.6 F 6.4
6:00 232.3 F 12.4
8:00 233.3 F 9.5
10:00 239.7 F 9.1
12:00 242.7 F 11.1
14:00 246.9 F 14.3
16:00 260.0 F 21.5
18:00 271.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.6 F 29.9
2:00 272.1 F 29.2
3:00 271.5 F 29.2
4:00 272.6 F 29.9

Min 271.4 F 28.9
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 204.9 F 0.0
10:00 223.4 F 0.0
15:00 230.5 F 0.0
20:00 228.4 F 0.0
25:00 224.4 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 18:40
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:52
CYCLE COMPLETE

10/06/2022 11:45 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____