
Midmark M11 - v1.0.5
Total Cycles: 280

Name: RITTER2
Sterilizer ID: V2313228
08/12/2021 06:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 87.6 F 0.0
4:00 79.8 F 0.1
6:00 105.7 F 0.3
8:00 134.2 F 3.0
10:00 161.6 F 5.9
12:00 188.8 F 9.3
14:00 200.8 F 11.6
16:00 215.1 F 6.8
18:00 228.3 F 10.9
20:00 229.3 F 7.5
22:00 241.7 F 12.4
24:00 239.5 F 10.5
26:00 251.5 F 16.5
28:00 261.9 F 22.6
30:00 271.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.6 F 29.5
2:00 272.4 F 29.9
3:00 271.7 F 29.1
4:00 272.4 F 29.7

Min 271.4 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.2 F 0.0
5:00 235.6 F 0.0
10:00 226.3 F 0.0
15:00 225.8 F 0.0
20:00 227.6 F 0.0
25:00 208.9 F 0.0
30:00 206.8 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 30:34
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:09:13
CYCLE COMPLETE

08/12/2021 08:13 AM
ACCEPTED: X
REJECTED:
Operator: U02