

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001352  
17:30 16/11/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:00  
PRESSURIZING 4:01  
STERILIZING 5:10  
136.4 C 224kPa 5:10  
136.4 C 225kPa 5:40  
136.4 C 225kPa 6:10  
136.4 C 224kPa 6:40  
136.2 C 223kPa 7:10  
136.3 C 224kPa 7:40  
136.4 C 224kPa 8:10  
136.3 C 224kPa 8:40  
136.3 C 224kPa 8:40  
VENTING 8:40  
AIR DRYING 9:34  
CYCLE COMPLETE 9:40  
-----

Digital Signature #  
5E231C74FA69EBD5  
-----