
Midmark M11 - v1.0.0
Total Cycles: 3223

Name: RITTER 1
Sterilizer ID: V2178198
05/12/2023 09:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.6 F 0.0
2:00 182.8 F 0.0
4:00 183.7 F 2.2
6:00 211.0 F 7.2
8:00 226.7 F 12.1
10:00 227.9 F 7.7
12:00 240.6 F 13.1
14:00 239.6 F 10.6
16:00 236.1 F 8.7
18:00 250.4 F 15.4
20:00 261.9 F 21.9
22:00 272.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.2 F 28.8
2:00 271.5 F 28.4
3:00 272.3 F 28.8
4:00 271.3 F 28.2

Min 271.2 F 28.1
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.6 F 0.0
5:00 234.8 F 0.0
10:00 232.5 F 0.0
15:00 228.7 F 0.0
20:00 227.7 F 0.0
25:00 229.2 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 22:27
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:57
CYCLE COMPLETE

05/12/2023 10:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____