
Midmark M11 - v1.0.0
Total Cycles: 3896

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2023 12:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.7 F 0.0
2:00 150.5 F 2.9
4:00 169.2 F 6.7
6:00 207.4 F 11.4
8:00 219.2 F 5.1
10:00 227.1 F 10.1
12:00 230.7 F 7.7
14:00 242.2 F 12.1
16:00 239.9 F 10.4
18:00 253.4 F 17.0
20:00 264.9 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.4 F 29.0
2:00 271.1 F 28.4
3:00 273.2 F 29.3
4:00 272.1 F 29.3

Min 270.9 F 28.2
Max 273.2 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 233.3 F 0.0
10:00 255.3 F 0.0
15:00 240.1 F 0.0
20:00 229.4 F 0.0
25:00 229.7 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:00:09
CYCLE COMPLETE

10/12/2023 01:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____