
Midmark M11 - v1.0.0
Total Cycles: 2807

Name: RITTER 1
Sterilizer ID: V2178198
01/23/2023 03:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.5 F 0.0
2:00 192.2 F 1.8
4:00 192.5 F 6.5
6:00 208.6 F 11.2
8:00 229.4 F 6.6
10:00 235.3 F 12.3
12:00 239.3 F 9.9
14:00 236.6 F 8.7
16:00 249.9 F 15.0
18:00 261.2 F 21.4
20:00 271.4 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 29.2
2:00 272.5 F 29.2
3:00 272.0 F 29.3
4:00 271.4 F 28.9

Min 270.7 F 28.1
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 247.7 F 0.0
10:00 227.1 F 0.0
15:00 230.7 F 0.0
20:00 228.7 F 0.0
25:00 228.1 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:31
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

01/23/2023 04:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____