
Midmark M11 - v1.0.0
Total Cycles: 961

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2021 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.4 F 0.0
2:00 187.7 F 3.8
4:00 202.0 F 8.1
6:00 225.3 F 13.4
8:00 229.4 F 8.1
10:00 240.4 F 13.4
12:00 239.4 F 10.9
14:00 237.6 F 9.2
16:00 251.1 F 15.9
18:00 262.6 F 22.5
20:00 272.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.4 F 28.8
2:00 272.2 F 29.0
3:00 272.1 F 28.5
4:00 272.7 F 29.2

Min 271.6 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.5 F 0.0
5:00 222.2 F 0.0
10:00 230.2 F 0.0
15:00 229.0 F 0.0
20:00 227.9 F 0.0
25:00 227.5 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

07/05/2021 12:58 PM
ACCEPTED: X
REJECTED:
Operator: U01