
Midmark M11 - v1.0.5
Total Cycles: 1230

Name: RITTER2
Sterilizer ID: V2313228
06/27/2022 03:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.8 F 0.0
2:00 139.4 F 2.2
4:00 155.2 F 5.6
6:00 185.1 F 10.0
8:00 210.5 F 6.0
10:00 221.6 F 9.3
12:00 232.1 F 9.6
14:00 239.1 F 11.8
16:00 238.6 F 10.4
18:00 252.1 F 17.1
20:00 263.8 F 24.1
22:00 271.4 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.2 F 30.1
2:00 272.7 F 29.8
3:00 271.7 F 29.4
4:00 271.2 F 29.1

Min 271.1 F 29.0
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 218.9 F 0.0
10:00 237.3 F 0.0
15:00 235.0 F 0.0
20:00 218.5 F 0.0
25:00 216.6 F 0.0
30:00 214.2 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:07
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

06/28/2022 06:43 AM
ACCEPTED: X
REJECTED:
Operator: U02