
Midmark M11 - v1.0.0
Total Cycles: 4819

Name: RITTER 1
Sterilizer ID: V2178198
08/14/2024 05:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.4 F 0.0
2:00 136.6 F 0.3
4:00 142.6 F 1.2
6:00 173.1 F 3.1
8:00 195.8 F 6.8
10:00 207.5 F 10.8
12:00 217.3 F 4.6
14:00 229.3 F 8.7
16:00 240.4 F 13.5
18:00 239.7 F 10.7
20:00 236.1 F 8.6
22:00 250.6 F 15.4
24:00 262.3 F 22.0
26:00 272.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.6 F 28.7
2:00 271.6 F 28.0
3:00 272.7 F 28.9
4:00 272.3 F 28.3

Min 271.6 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 195.1 F 0.0
10:00 189.6 F 0.0
15:00 190.4 F 0.0
20:00 192.5 F 0.0
25:00 195.1 F 0.0
30:00 195.5 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:22
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:04:47
CYCLE COMPLETE

08/14/2024 06:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____