
Midmark M11 - v1.0.5
Total Cycles: 4177

Name: RITTER2
Sterilizer ID: V2313228
12/03/2024 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.2 F 0.0
2:00 157.2 F 0.5
4:00 169.7 F 2.8
6:00 196.5 F 6.8
8:00 208.3 F 11.3
10:00 221.7 F 6.2
12:00 232.7 F 11.1
14:00 233.0 F 8.3
16:00 241.9 F 11.3
18:00 245.7 F 13.6
20:00 257.9 F 20.1
22:00 268.7 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.6 F 29.7
2:00 271.7 F 29.0
3:00 271.8 F 29.7
4:00 272.4 F 29.5

Min 271.3 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 214.3 F 0.0
10:00 228.9 F 0.0
15:00 225.8 F 0.0
20:00 223.9 F 0.0
25:00 217.6 F 0.0
30:00 200.4 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:12
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:01:60
CYCLE COMPLETE

12/03/2024 02:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____