
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006035
9:59 24/07/2023

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 3:46
STERILIZING 4:47
136.5 C 224kPa 4:47
135.8 C 219kPa 5:17
136.3 C 224kPa 5:47
136.2 C 222kPa 6:17
136.3 C 222kPa 6:47
136.2 C 223kPa 7:17
136.2 C 222kPa 7:47
136.3 C 220kPa 8:17
136.3 C 220kPa 8:17
VENTING 8:17
AIR DRYING 9:12
CYCLE COMPLETE 20:19

Digital Signature #
C73909A3599A1E53
