
Midmark M11 - v1.0.5
Total Cycles: 502

Name: RITTER2
Sterilizer ID: V2313228
10/29/2021 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.1 F 0.0
2:00 99.7 F 0.6
4:00 119.1 F 2.6
6:00 150.5 F 5.7
8:00 172.0 F 9.8
10:00 194.0 F 6.5
12:00 212.1 F 6.7
14:00 225.5 F 11.2
16:00 228.6 F 7.4
18:00 241.6 F 13.1
20:00 241.3 F 11.6
22:00 253.5 F 17.8
24:00 264.6 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.5 F 29.7
2:00 272.0 F 29.7
3:00 272.2 F 29.5
4:00 272.1 F 29.8

Min 271.6 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.7 F 0.0
5:00 201.1 F 0.0
10:00 204.9 F 0.0
15:00 205.5 F 0.0
20:00 204.8 F 0.0
25:00 201.5 F 0.0
30:00 199.2 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 25:53
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:04:01
CYCLE COMPLETE

10/29/2021 09:59 AM
ACCEPTED: X
REJECTED:
Operator: U02