
Midmark M11 - v1.0.0
Total Cycles: 2084

Name: RITTER 1
Sterilizer ID: V2178198
07/04/2022 03:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.9 F 0.0
2:00 183.2 F 3.9
4:00 196.2 F 8.4
6:00 217.1 F 13.4
8:00 226.7 F 7.9
10:00 238.7 F 13.3
12:00 237.7 F 10.4
14:00 237.5 F 8.3
16:00 250.1 F 15.3
18:00 261.9 F 22.2
20:00 272.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.6 F 29.2
2:00 271.9 F 28.5
3:00 272.1 F 29.1
4:00 272.4 F 29.0

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 212.9 F 0.0
10:00 227.4 F 0.0
15:00 235.6 F 0.0
20:00 242.5 F 0.0
25:00 227.8 F 0.0
30:00 219.9 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 20:26
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

07/04/2022 04:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____