

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2934

Name: RITTER2  
Sterilizer ID: V2313228  
09/26/2023 04:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.4 F 0.0  
2:00 177.3 F 3.8  
4:00 199.9 F 9.1  
6:00 216.8 F 6.8  
8:00 230.2 F 10.6  
10:00 233.5 F 8.1  
12:00 239.9 F 10.2  
14:00 248.5 F 14.6  
16:00 261.5 F 22.1  
18:00 272.6 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.6  
1:00 271.9 F 28.8  
2:00 272.3 F 29.3  
3:00 271.3 F 28.8  
4:00 271.7 F 28.8  
  
Min 271.1 F 28.4  
Max 272.5 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.6 F 0.0  
5:00 209.9 F 0.0  
10:00 198.9 F 0.0  
15:00 198.2 F 0.0  
20:00 196.2 F 0.0  
25:00 194.2 F 0.0  
30:00 192.3 F 0.0  
Drying Complete

FILLING: 2:17  
HEATING: 18:12  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:57:36  
CYCLE COMPLETE

09/26/2023 05:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_