
Midmark M11 - v1.0.5
Total Cycles: 4323

Name: RITTER2
Sterilizer ID: V2313228
01/31/2025 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.2 F 0.0
2:00 133.7 F 1.1
4:00 142.1 F 3.4
6:00 164.6 F 6.8
8:00 185.3 F 10.5
10:00 204.4 F 5.6
12:00 217.2 F 6.8
14:00 223.3 F 10.9
16:00 230.1 F 7.4
18:00 238.5 F 11.9
20:00 237.2 F 10.0
22:00 248.7 F 15.7
24:00 259.2 F 21.5
26:00 269.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.7
1:00 272.7 F 30.3
2:00 271.8 F 29.6
3:00 272.4 F 30.4
4:00 272.4 F 30.0

Min 271.5 F 29.6
Max 272.8 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 213.8 F 0.0
10:00 212.9 F 0.0
15:00 203.9 F 0.0
20:00 200.3 F 0.0
25:00 219.7 F 0.0
30:00 221.4 F 0.0

Drying Complete

FILLING: 0:57
HEATING: 27:12
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:06:02
CYCLE COMPLETE

01/31/2025 01:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____