

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3278

Name: RITTER2  
Sterilizer ID: V2313228  
01/04/2024 02:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 165.4 F 0.0  
2:00 174.0 F 3.4  
4:00 201.0 F 8.3  
6:00 216.7 F 10.2  
8:00 228.7 F 9.3  
10:00 235.5 F 9.4  
12:00 244.8 F 13.0  
14:00 245.4 F 13.2  
16:00 258.2 F 20.2  
18:00 269.8 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.9  
1:00 272.7 F 29.7  
2:00 272.0 F 29.1  
3:00 271.2 F 28.9  
4:00 272.3 F 29.7  
  
Min 271.1 F 28.7  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.0 F 0.0  
5:00 227.7 F 0.0  
10:00 211.5 F 0.0  
15:00 196.5 F 0.0  
20:00 192.9 F 0.0  
25:00 190.4 F 0.0  
30:00 188.8 F 0.0  
Drying Complete

FILLING: 1:54  
HEATING: 19:00  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 00:58:22  
CYCLE COMPLETE

01/04/2024 03:15 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_