
Midmark M11 - v1.0.5
Total Cycles: 822

Name: RITTER2
Sterilizer ID: V2313228
02/22/2022 01:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.7 F 0.0
2:00 164.2 F 2.1
4:00 175.6 F 6.3
6:00 190.4 F 11.6
8:00 223.4 F 6.6
10:00 227.6 F 12.8
12:00 238.1 F 10.4
14:00 234.4 F 8.7
16:00 249.7 F 15.8
18:00 261.6 F 22.9
20:00 272.5 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 30.0
1:00 272.7 F 30.9
2:00 273.4 F 30.6
3:00 272.3 F 30.1
4:00 271.7 F 30.1

Min 271.5 F 29.9
Max 273.9 F 31.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.6 F 0.0
5:00 230.7 F 0.0
10:00 248.2 F 0.0
15:00 239.4 F 0.0
20:00 227.4 F 0.0
25:00 230.7 F 0.0
30:00 230.9 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 21:17
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:00:08
CYCLE COMPLETE

02/22/2022 02:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____