
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000475
14:30 12/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:10
STERILIZING 4:19
136.0 C 228kPa 4:19
136.0 C 229kPa 4:49
136.4 C 233kPa 5:19
136.3 C 233kPa 5:49
136.4 C 233kPa 6:19
136.3 C 232kPa 6:49
136.3 C 233kPa 7:19
136.2 C 230kPa 7:49
136.2 C 230kPa 7:49
VENTING 7:49
AIR DRYING 8:43
CYCLE COMPLETE 10:33

Digital Signature #
A6248E34F2205933
