

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 008252  
10:15 28/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:15  
PRESSURIZING 3:16  
STERILIZING 4:22  
136.4 C 224kPa 4:22  
136.3 C 224kPa 4:52  
136.4 C 225kPa 5:22  
136.4 C 223kPa 5:52  
136.4 C 223kPa 6:22  
136.4 C 224kPa 6:52  
136.3 C 224kPa 7:22  
136.4 C 223kPa 7:52  
136.4 C 223kPa 7:52  
VENTING 7:52  
AIR DRYING 8:44  
CYCLE COMPLETE 9:20  
-----

Digital Signature #  
C7D3D90B9E2295AB  
-----