

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4899

Name: RITTER 1  
Sterilizer ID: V2178198  
09/12/2024 11:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.5 F 0.0  
2:00 151.9 F 0.1  
4:00 161.1 F 0.4  
6:00 201.6 F 1.2  
8:00 207.9 F 5.2  
10:00 221.4 F 9.1  
12:00 233.4 F 13.3  
14:00 233.3 F 9.2  
16:00 241.6 F 10.7  
18:00 243.3 F 12.1  
20:00 241.0 F 10.6  
22:00 253.5 F 16.8  
24:00 263.9 F 22.9  
26:00 272.6 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 271.4 F 28.1  
2:00 272.3 F 28.7  
3:00 272.4 F 28.5  
4:00 271.8 F 28.2

Min 271.1 F 27.8  
Max 272.6 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 214.1 F 0.0  
10:00 197.7 F 0.0  
15:00 217.0 F 0.0  
20:00 229.2 F 0.0  
25:00 238.9 F 0.0  
30:00 233.5 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 26:14  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:04:45  
CYCLE COMPLETE

09/12/2024 12:41 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_