
Midmark M11 - v1.0.5
Total Cycles: 3703

Name: RITTER2
Sterilizer ID: V2313228
06/14/2024 03:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.5 F 0.0
2:00 173.6 F 3.2
4:00 196.5 F 8.2
6:00 214.1 F 12.4
8:00 226.6 F 9.0
10:00 234.3 F 10.1
12:00 242.0 F 13.0
14:00 242.5 F 12.2
16:00 255.7 F 19.1
18:00 267.0 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.4 F 29.7
2:00 271.5 F 29.4
3:00 272.3 F 29.6
4:00 271.4 F 29.2

Min 271.3 F 29.0
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 214.0 F 0.0
10:00 218.9 F 0.0
15:00 219.2 F 0.0
20:00 218.9 F 0.0
25:00 218.0 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 19:25
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

06/14/2024 04:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____