

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4376

Name: RITTER2  
Sterilizer ID: V2313228  
02/24/2025 04:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.1 F 0.0  
2:00 167.3 F 3.1  
4:00 193.4 F 8.3  
6:00 217.8 F 8.8  
8:00 231.0 F 9.2  
10:00 234.6 F 9.5  
12:00 244.5 F 13.4  
14:00 244.1 F 12.8  
16:00 257.9 F 20.2  
18:00 270.0 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.1  
1:00 272.1 F 29.1  
2:00 272.8 F 29.6  
3:00 272.6 F 29.8  
4:00 272.0 F 29.0  
  
Min 271.8 F 28.9  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.0 F 0.0  
5:00 186.6 F 0.0  
10:00 201.9 F 0.0  
15:00 210.5 F 0.0  
20:00 212.5 F 0.0  
25:00 211.4 F 0.0  
30:00 207.5 F 0.0  
  
Drying Complete

FILLING: 2:03  
HEATING: 18:50  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 00:58:14  
CYCLE COMPLETE

02/24/2025 05:46 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_