
Midmark M11 - v1.0.5
Total Cycles: 4129

Name: RITTER2
Sterilizer ID: V2313228
11/18/2024 10:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes

C102: FILL MODE

** ITEMS NOT STERILE **

11/18/2024 10:53 AM