

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001734  
12:03 15/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:03  
PRESSURIZING 2:08  
STERILIZING 3:06  
136.0 C 220kPa 3:06  
136.2 C 224kPa 3:36  
136.4 C 225kPa 4:06  
136.0 C 222kPa 4:36  
136.2 C 224kPa 5:06  
136.4 C 224kPa 5:36  
136.4 C 225kPa 6:06  
136.4 C 225kPa 6:36  
136.4 C 225kPa 6:36  
VENTING 6:37  
AIR DRYING 7:26  
CYCLE COMPLETE 7:34  
-----

Digital Signature #  
25FDB07F79B974E2  
-----