

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4071

Name: RITTER2  
Sterilizer ID: V2313228  
10/31/2024 11:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 94.9 F 0.0  
2:00 94.8 F 0.0  
4:00 109.2 F 0.0  
6:00 147.4 F 0.1  
8:00 176.7 F 0.8  
10:00 194.8 F 4.0  
12:00 203.4 F 7.7  
14:00 215.9 F 11.4  
16:00 222.3 F 5.8  
18:00 233.7 F 10.7  
20:00 233.6 F 8.4  
22:00 244.2 F 13.5  
24:00 243.3 F 12.4  
26:00 254.9 F 18.4  
28:00 265.2 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.0  
1:00 272.0 F 29.6  
2:00 272.2 F 29.3  
3:00 271.5 F 29.2  
4:00 272.5 F 29.7

Min 271.3 F 28.9  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.6 F 0.0  
5:00 200.8 F 0.0  
10:00 215.5 F 0.0  
15:00 230.3 F 0.0  
20:00 220.3 F 0.0  
25:00 217.4 F 0.0  
30:00 226.4 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 29:51  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:08:33  
CYCLE COMPLETE

10/31/2024 12:33 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_