
Midmark M11 - v1.0.0
Total Cycles: 3103

Name: RITTER 1
Sterilizer ID: V2178198
04/13/2023 07:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.2 F 0.0
2:00 81.6 F 0.0
4:00 110.2 F 0.0
6:00 132.9 F 0.0
8:00 155.5 F 1.2
10:00 182.8 F 4.4
12:00 199.0 F 7.8
14:00 212.3 F 11.8
16:00 218.6 F 4.9
18:00 228.3 F 9.0
20:00 239.1 F 13.3
22:00 238.0 F 10.1
24:00 236.6 F 8.6
26:00 247.9 F 14.1
28:00 259.3 F 20.4
30:00 269.9 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.5 F 28.8
2:00 271.6 F 28.4
3:00 272.6 F 28.8
4:00 271.5 F 28.3

Min 271.5 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 207.0 F 0.0
10:00 217.8 F 0.0
15:00 223.5 F 0.0
20:00 218.9 F 0.0
25:00 219.5 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 30:56
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:08
CYCLE COMPLETE

04/13/2023 08:20 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____