
Midmark M11 - v1.0.5
Total Cycles: 4307

Name: RITTER2
Sterilizer ID: V2313228
01/28/2025 05:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.6 F 0.0
2:00 111.0 F 0.8
4:00 124.3 F 3.2
6:00 146.7 F 6.4
8:00 172.2 F 10.2
10:00 198.4 F 7.0
12:00 215.7 F 6.2
14:00 225.0 F 10.1
16:00 233.1 F 10.0
18:00 235.9 F 11.1
20:00 235.5 F 9.1
22:00 247.9 F 15.0
24:00 258.4 F 21.0
26:00 268.1 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.7
1:00 272.1 F 30.3
2:00 271.2 F 29.7
3:00 272.0 F 30.2
4:00 271.0 F 29.7

Min 271.0 F 29.6
Max 272.7 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 217.2 F 0.0
10:00 218.2 F 0.0
15:00 191.0 F 0.0
20:00 183.6 F 0.0
25:00 182.4 F 0.0
30:00 182.3 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 27:10
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:06:08
CYCLE COMPLETE

01/28/2025 06:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____