

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001694  
12:46 05/02/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:36  
PRESSURIZING 2:41  
STERILIZING 3:43  
136.4 C 224kPa 3:43  
136.4 C 226kPa 4:13  
136.3 C 225kPa 4:43  
136.4 C 225kPa 5:13  
136.3 C 225kPa 5:43  
136.3 C 224kPa 6:13  
136.4 C 225kPa 6:43  
136.3 C 225kPa 7:13  
136.3 C 225kPa 7:13  
VENTING 7:13  
AIR DRYING 8:06  
CYCLE COMPLETE 8:14  
-----

Digital Signature #  
953C3DADE87524DA  
-----