
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002180
10:53 29/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:15
PRESSURIZING 2:18
STERILIZING 3:16
136.4 C 225kPa 3:16
136.4 C 225kPa 3:46
136.2 C 224kPa 4:16
136.3 C 225kPa 4:46
136.4 C 226kPa 5:16
136.4 C 225kPa 5:46
136.3 C 223kPa 6:16
136.1 C 223kPa 6:46
136.1 C 223kPa 6:46
VENTING 6:46
AIR DRYING 7:37
CYCLE COMPLETE 21:15

Digital Signature #
71E776FB7C742E35
