
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002425
7:19 13/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:24
PRESSURIZING 1:27
STERILIZING 2:34
136.4 C 220kPa 2:34
135.9 C 220kPa 3:04
136.0 C 217kPa 3:34
136.2 C 220kPa 4:04
136.4 C 221kPa 4:34
136.4 C 220kPa 5:04
136.4 C 223kPa 5:34
136.4 C 221kPa 6:04
136.4 C 221kPa 6:04
VENTING 6:04
AIR DRYING 6:56
CYCLE COMPLETE 7:05

Digital Signature #
1BC1D90B937F10F2
