
Midmark M11 - v1.0.0
Total Cycles: 3976

Name: RITTER 1
Sterilizer ID: V2178198
11/03/2023 08:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 73.5 F 0.3
4:00 100.0 F 1.2
6:00 126.6 F 3.5
8:00 148.3 F 6.2
10:00 170.4 F 9.4
12:00 188.9 F 12.9
14:00 202.6 F 3.2
16:00 213.9 F 6.3
18:00 224.2 F 9.7
20:00 232.5 F 11.0
22:00 235.5 F 9.5
24:00 239.9 F 10.3
26:00 245.0 F 12.8
28:00 256.6 F 18.8
30:00 267.0 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.6 F 29.0
2:00 271.6 F 28.2
3:00 272.7 F 29.0
4:00 271.7 F 28.2

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.1 F 0.0
5:00 184.9 F 0.0
10:00 217.4 F 0.0
15:00 226.8 F 0.0
20:00 222.7 F 0.0
25:00 220.5 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:23
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:09:42
CYCLE COMPLETE

11/03/2023 09:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____