

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4205

Name: RITTER 1  
Sterilizer ID: V2178198  
01/15/2024 02:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 112.1 F 0.0  
2:00 113.3 F 1.2  
4:00 182.1 F 3.4  
6:00 165.9 F 6.9  
8:00 188.7 F 10.8  
10:00 215.2 F 3.8  
12:00 219.1 F 6.6  
14:00 229.1 F 10.6  
16:00 232.0 F 7.5  
18:00 240.5 F 11.1  
20:00 235.8 F 8.6  
22:00 249.2 F 14.7  
24:00 260.2 F 20.8  
26:00 270.3 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.4 F 29.1  
2:00 271.4 F 28.4  
3:00 272.1 F 28.8  
4:00 272.3 F 29.1

Min 270.7 F 27.9  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.1 F 0.0  
5:00 234.7 F 0.0  
10:00 205.7 F 0.0  
15:00 202.1 F 0.0  
20:00 231.4 F 0.0  
25:00 226.8 F 0.0  
30:00 226.8 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 26:47  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:05:23  
CYCLE COMPLETE

01/15/2024 03:41 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_