
Midmark M11 - v1.0.0
Total Cycles: 3988

Name: RITTER 1
Sterilizer ID: V2178198
11/07/2023 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.3 F 0.0
2:00 169.8 F 2.0
4:00 171.2 F 5.3
6:00 197.1 F 9.5
8:00 219.7 F 6.5
10:00 222.2 F 8.2
12:00 233.3 F 12.9
14:00 235.6 F 9.8
16:00 238.8 F 9.2
18:00 247.3 F 14.2
20:00 258.9 F 20.5
22:00 269.6 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.7 F 29.3
2:00 271.7 F 28.7
3:00 271.6 F 29.0
4:00 272.7 F 29.3

Min 271.2 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 220.5 F 0.0
10:00 223.1 F 0.0
15:00 225.5 F 0.0
20:00 226.1 F 0.0
25:00 229.1 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 23:03
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:02:08
CYCLE COMPLETE

11/07/2023 11:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____