
Midmark M11 - v1.0.5
Total Cycles: 2767

Name: RITTER2
Sterilizer ID: V2313228
08/15/2023 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.7 F 0.0
2:00 163.0 F 0.0
4:00 181.3 F 0.0
6:00 205.1 F 1.8
8:00 219.7 F 6.2
10:00 233.8 F 10.9
12:00 232.2 F 8.1
14:00 242.1 F 11.7
16:00 244.8 F 13.1
18:00 244.3 F 12.6
20:00 257.1 F 19.4
22:00 268.5 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.0 F 28.8
2:00 272.6 F 29.3
3:00 271.8 F 29.1
4:00 271.8 F 28.7

Min 271.5 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 205.0 F 0.0
10:00 205.6 F 0.0
15:00 206.4 F 0.0
20:00 205.0 F 0.0
25:00 204.4 F 0.0
30:00 207.1 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 23:07
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:01:55
CYCLE COMPLETE

08/15/2023 10:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____