

-----  
Midmark M11 - v1.0.0  
Total Cycles: 875

Name: RITTER 1  
Sterilizer ID: V2178198  
06/09/2021 06:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.6 F 0.0  
2:00 75.3 F 0.0  
4:00 82.4 F 0.1  
6:00 105.5 F 0.1  
8:00 140.7 F 0.1  
10:00 172.1 F 0.2  
12:00 193.7 F 1.6  
14:00 205.7 F 4.6  
16:00 213.4 F 7.3  
18:00 224.4 F 10.5  
20:00 234.4 F 8.8  
22:00 234.4 F 9.9  
24:00 238.6 F 9.9  
26:00 244.1 F 12.4  
28:00 241.2 F 10.8  
30:00 253.0 F 16.7  
32:00 263.9 F 23.1  
34:00 271.5 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.5 F 28.7  
2:00 271.5 F 28.2  
3:00 272.4 F 28.7  
4:00 271.5 F 28.1

Min 271.4 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.1 F 0.0  
5:00 213.4 F 0.0  
10:00 221.8 F 0.0  
15:00 214.5 F 0.0  
20:00 219.6 F 0.0  
25:00 219.7 F 0.0  
30:00 221.1 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 34:05  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:12:14  
CYCLE COMPLETE

06/09/2021 07:56 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_