
Midmark M11 - v1.0.0
Total Cycles: 2505

Name: RITTER 1
Sterilizer ID: V2178198
11/02/2022 05:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.7 F 0.0
2:00 168.4 F 3.6
4:00 197.8 F 7.7
6:00 207.8 F 12.8
8:00 226.1 F 6.8
10:00 237.6 F 12.1
12:00 235.2 F 9.0
14:00 240.5 F 9.7
16:00 247.1 F 13.6
18:00 259.2 F 20.3
20:00 270.2 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 28.9
2:00 271.4 F 28.3
3:00 272.5 F 29.1
4:00 271.9 F 28.5

Min 271.4 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 237.8 F 0.0
10:00 234.2 F 0.0
15:00 227.4 F 0.0
20:00 228.5 F 0.0
25:00 230.8 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 20:52
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:15
CYCLE COMPLETE

11/02/2022 06:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____