
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001999
9:40 10/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:10
STERILIZING 3:09
136.0 C 220kPa 3:09
136.3 C 225kPa 3:39
136.4 C 226kPa 4:09
136.3 C 225kPa 4:39
136.3 C 224kPa 5:09
136.2 C 225kPa 5:39
136.2 C 224kPa 6:09
136.4 C 226kPa 6:39
136.4 C 226kPa 6:39
VENTING 6:39
AIR DRYING 7:30
CYCLE COMPLETE 19:42

Digital Signature #
898E403FC448E195
