
Midmark M11 - v1.0.5
Total Cycles: 1592

Name: RITTER2
Sterilizer ID: V2313228
10/13/2022 09:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.2 F 0.0
2:00 190.4 F 3.1
4:00 207.2 F 8.1
6:00 216.6 F 10.4
8:00 229.5 F 10.4
10:00 231.9 F 7.8
12:00 239.7 F 11.0
14:00 245.3 F 13.8
16:00 257.8 F 20.6
18:00 269.2 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.4 F 29.8
2:00 271.5 F 29.8
3:00 272.1 F 29.7
4:00 271.4 F 29.6

Min 271.3 F 29.3
Max 272.5 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 219.6 F 0.0
10:00 229.6 F 0.0
15:00 236.9 F 0.0
20:00 226.7 F 0.0
25:00 221.3 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 2:39
HEATING: 19:01
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

10/13/2022 10:17 AM
ACCEPTED: X
REJECTED:
Operator: U02