
Midmark M11 - v1.0.0
Total Cycles: 4166

Name: RITTER 1
Sterilizer ID: V2178198
12/28/2023 04:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.0 F 0.0
2:00 135.8 F 1.5
4:00 158.4 F 4.5
6:00 195.1 F 9.0
8:00 208.0 F 9.8
10:00 220.1 F 6.6
12:00 232.2 F 11.0
14:00 230.9 F 6.8
16:00 243.7 F 12.5
18:00 240.2 F 10.5
20:00 252.9 F 16.7
22:00 264.0 F 23.2
24:00 271.1 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.3 F 28.7
2:00 272.7 F 29.1
3:00 271.9 F 29.0
4:00 272.0 F 28.6

Min 271.5 F 28.5
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.1 F 0.0
5:00 239.9 F 0.0
10:00 231.8 F 0.0
15:00 227.0 F 0.0
20:00 226.6 F 0.0
25:00 214.7 F 0.0
30:00 205.0 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 24:44
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:03:28
CYCLE COMPLETE

12/28/2023 05:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____