

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2558

Name: RITTER 1  
Sterilizer ID: V2178198  
11/15/2022 09:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.0 F 0.0  
2:00 151.4 F 1.7  
4:00 165.0 F 4.9  
6:00 187.9 F 9.5  
8:00 206.3 F 6.1  
10:00 224.5 F 9.0  
12:00 234.0 F 10.1  
14:00 240.0 F 11.1  
16:00 236.1 F 8.7  
18:00 250.6 F 15.6  
20:00 262.4 F 22.5  
22:00 272.6 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.6  
1:00 272.1 F 29.3  
2:00 271.4 F 28.5  
3:00 271.8 F 29.3  
4:00 272.0 F 29.1

Min 271.1 F 28.5  
Max 272.1 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.0 F 0.0  
5:00 215.6 F 0.0  
10:00 224.9 F 0.0  
15:00 233.4 F 0.0  
20:00 231.5 F 0.0  
25:00 226.3 F 0.0  
30:00 224.9 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 22:17  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:30  
CYCLE COMPLETE

11/15/2022 10:34 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_