

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000238  
8:34 29/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:10  
PRESSURIZING 4:12  
STERILIZING 5:48  
136.0 C 229kPa 5:48  
136.4 C 232kPa 6:18  
136.4 C 233kPa 6:48  
136.4 C 233kPa 7:18  
136.2 C 231kPa 7:48  
136.4 C 233kPa 8:18  
136.4 C 233kPa 8:48  
136.3 C 232kPa 9:18  
136.3 C 232kPa 9:18  
VENTING 9:18  
AIR DRYING 10:21  
CYCLE COMPLETE 25:21  
-----

Digital Signature #  
FFBA98D9D2191DBA  
-----