
Midmark M11 - v1.0.5
Total Cycles: 1502

Name: RITTER2
Sterilizer ID: V2313228
09/19/2022 12:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.9 F 0.0
2:00 195.5 F 4.1
4:00 204.7 F 9.6
6:00 223.7 F 7.3
8:00 231.2 F 11.4
10:00 234.9 F 9.5
12:00 237.2 F 9.3
14:00 249.8 F 15.5
16:00 262.2 F 22.5
18:00 272.9 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.9 F 30.1
2:00 272.1 F 29.4
3:00 272.2 F 29.9
4:00 272.9 F 30.0

Min 271.6 F 29.3
Max 273.3 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 233.1 F 0.0
10:00 231.2 F 0.0
15:00 229.9 F 0.0
20:00 229.5 F 0.0
25:00 230.5 F 0.0
30:00 230.4 F 0.0

Drying Complete

FILLING: 2:32
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

09/19/2022 01:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____